## EN 1427: Determination of softening point (T<sub>R&B</sub>) **Ring and ball method**

 $\checkmark$ 

## $28 \text{ °C} < T_{R\&B} < 150 \text{ °C}$ :

- Two brass rings and two steel balls
- Controlled rate heating
- Falling distance 25 ± 0,4 mm
- $\succ$ Ring holder and assembly – Upper edge of the rings ≈ 50 mm below liquid level
- Glass beaker D > 85 mm , H > 120 mm

## Temperature measuring device

- Distilled bath water  $(T = 0 90 \circ C)$
- Glycerol (T = 30 155 °C)
- Magnetic stirrer placed at the bottom of the beaker  $\approx$  40 mm and diameter 8 mm, 100 r/ min.

## **Bath liquid**

- Distilled bath water (28 °C < T<sub>R&B</sub> < 80 °C )
- Solution  $(80 \text{ °C} < T_{R\&B} < 150 \text{ °C})$
- **Release agent** 
  - Vaseline or (glycerol + mineral talc)
  - Pouring plate (≈ 50 mm x 75 mm and 1,5 mm thick)
- Prepare the samples according to EN 12594
- Heat the two rings to a temperature < 100 °C above the expected T<sub>R&B</sub>
- Place the rings in the pouring plate treated with the release  $\checkmark$ agent
- $\checkmark$ Pour slightly the bitumen in the rings
- Let the specimens cool down at least 30 min.
- Cut the excess of binder with a hot knife before testing to ensure horizontal surface of the specimen



- Record the  $T_{R\&B}$  when the ball interrupts the ray of light of the automatic equipment.
- If the difference between the two  $T_{R\&B}$  exceeds 1 °C for  $T_{R\&B}$  < 80 °C and 2 °C for  $T_{R\&B} > 80$  °C repeat the test
- Express the average result of the two T<sub>R&B</sub> measurements for



- Take all the fixtures outside the bath and dry them gently Place all the fixtures in the
  - - 5 °C/min. The first 3 min. are intended to settle the heating rate of 5 °C/min.

from below with a rate

Initial T for bath liquid:

- - Water: 5 ± 1 °C
  - Glycerol: 30 ± 1°C
- Place the apparatus, the ring holders, the steel balls and the sample rings in the appropriate T in the bath inside a thermostatic device e.g. a refrigerator, for 15-20 min.

appropriate position as quickly as possible and the balls using forceps in the ball centering guides Start the test

The temperature increase after the first 3 min. should be 4,4–5,6 °C/min.

For the automatic equipment

the starting T should be

ensured that is within the

range of  $5 \pm 1$  or  $30 \pm 1$  °C

Stir the bath liquid and heat

- $T_{R\&B} < 80 \text{ °C in the}$ nearest 0,2 °C
- T<sub>R&B</sub> > 80 °C in the nearest 0,5 °C



Contact	
Lab manager	Johan Blom
Device manager	Georgios Pipintakos

